



UNIVERSITY OF DELHI - SOUTH CAMPUS

Department of Genetics
Benito Juarez Road
Dhaura Kuan
New Delhi-110 021 India

Tel : 91-11-24112503/(Ext. 172)
Fax : 91-11-24112761
E-mail : udscgenetics@gmail.com

Ref. SDC/Genetics /

Specification for Yeast Tetrad Dissection Microscope System

Please quote the lowest price for the equipment with the following specification within 3 weeks of the above notification on the Delhi University website.

Price should be inclusive of Duty paid, custom clearance and transportation to University of Delhi in case quotation is in foreign currency

S.No.	Technical Specification
1	Microscope with Provision for 5x and 10x objective lens and a 10x eye piece or better
2	The Microscope stage should be modified to hold petridishes with 90mm diameter
3	The microscope stage should also have graduated markings' on both X and Yaxis. Click stops should be provided at every 5 mm interval on both X and Yaxis.
4	Microscope should have a halogen lamp to illuminate through the agar media in the petridish. A day light filter is preferred to reduce eye strain.
5	A micromanipulator with joystick should be mounted on the microscope for making fine movements of the attached needle.
6	A dissection needle kit with spare needles should be provided with the microscope

• Eligibility Criteria


- Bidder should be Manufacturer/ Authorized Partner/ Reseller of the manufacturer and a Letter of Authorization from manufacturer for the same and specific to the tender should also be enclosed. The bidder should also be the Authorized Service Provider.
- The Bidder shall provide the Registration number of the firm along with the LST/ CST/ WCT No. and the PAN Number issued by the concerned authorities.
- The bidder must have experience of supplying similar orders with at least 3 orders of similar level. The bidder must also enclose documentary evidence of supplying and installation within the last 3 financial years for any organization preferably from Educational and R & D organization.

• No Commitment to Accept Lowest or Any Tender

- University of Delhi shall be under no obligation to accept the lowest or any other offer received in response to this tender notice and shall be entitled to reject any or all offers. University of Delhi will not be obliged to meet and have discussions with any vendor, and or to listen to any representations.

• Warranty /Comprehensive Maintenance

- The vendor shall undertake comprehensive support for at least 3(three) years. This would cover the replacement/repair of supplied items and accessories supplied by vendor at the place of installation and also complete installation and satisfactory operation of the supplied facility.


Kavitha Datta, Ph.D.
Assistant Professor
Department of Genetics
University of Delhi South Campus
New Delhi-110021