

UNIVERSITY OF DELHI SOUTH CAMPUS
Centre for Innovation in Infectious Disease Research, Education and Training (CIIDRET)

First Floor Engineering Department
New Delhi-110021
Tel.: 91-11-24115883
E-mail: vkchaudhary@south.du.ac.in

E-tender Notice

Ref. No. : UDSC/CIIDRET/VKC/2018/001


March 26, 2018

Tenders are hereby invited in two-bid system from manufactures or their authorized dealers for the **supply of Chest Deep Freezer [Quantity: Three (03) numbers]** with the specification given below:

1. The Deep Freezer should have 2 insulated doors with independent locking system.
2. The Deep Freezer should have digital display and temperature controller with features to set differential.
3. The temperature should go up to -25°C.
4. The Net capacity should be ~ 380 liters.
5. Refrigerant: R-134 A
6. External Dimensions in inches (Width: 57.5, Depth: 26.5, Height: 33)
7. Internal Dimensions in inches (Width: 52.2, Depth: 19.9, usable Height in main section: 25.5)
8. Inner cabinet should be made of Pre-Painted steel
9. Wheels and Door Lockable option
10. Power Supply: 220V to 240V - 50HZ
11. Power Consumption: 210 Watts
12. PUF insulation of 60 mm thickness
13. The compressor should carry 3-years warranty
14. Should come with one basket.

Terms and Conditions:

1. The bids should be uploaded on to the e-procurement website (<https://eprocure.gov.in>) within 21 days of the date of the advertisement (latest by 16th April 2018). The Bidders/Vendors are advised to follow the instructions provided in the 'Instructions for Online Bid Submission' for the e-submission of the bids online through the Central Public Procurement Portal for e-Procurement at <https://eprocure.gov.in/eprocure/app/>.
2. Quotations have to be submitted in two bid systems. The First part 'Technical bid', should consists of all technical details and supporting documents with terms and conditions. A compliance sheet (in the attached format; Annexure I) must be filled by the vendor against each point and giving reference of the same (page no., line no.) in the supporting company brochure/document.
3. The second part 'Financial bid', should contain item-wise pricing of items mentioned in the technical bid. The Financial Quotations should contain price of the equipment, discount if any, packaging and forwarding charges, Air Freight and insurance charges. The price quoted should be F.O.R destination price in INR against customs duty exemption certificates provided by the buyer and inclusive of standard installation.
4. Both the quotation/bids are to be sealed in separate cover duly super scribed by the supplier and then placed in a bigger cover, which should also be sealed and super scribed.
5. The supplier should also provide a list of at least five users (name, phone number and email addresses) in academic Institutions for the item quote in the bid. Also, provide certificate from the existing users (at least three) about the performance of the machine, quality of services and other support.
6. Payment will be made by RTGS mode within 30 days of satisfactory installation of the Freezer.
7. The quote should be valid for 90 days from the due date.
8. Authorization certificate from the manufacturer should be attached.
9. The institute reserves the right to order equipment with better quality and suitability over lower price and to accept or reject any or all quotations without assigning reasons thereof. The tender may also be cancelled without providing reasons.


Professor Vijay K. Chaudhary
निदेशक/ Director
नवपरिवर्तन केन्द्र - CIIDRET, 3rd फ्लोर, 3rd फ्लोर और प्रशिक्षण
Center for Innovation in Infectious Disease
Research, Education & Training (CIIDRET)
दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - ११००२१
University of Delhi South Campus, New Delhi-110021


Annexure I

Sheet for technical compliance for Chest type Deep Freezer

S. No	Feature	% Compliance (describe deviation)	Reference of the specification (page no., line no. in the product catalogue)
1.	The Deep Freezer should have 2 insulated doors with independent locking system.		
2.	The Deep Freezer should have digital display and temperature controller with features to set differential.		
3.	The temperature should go up to -25°C.		
4.	The Net capacity should be ~ 380 litres.		
5.	Refrigerant: R-134 A		
6.	External Dimensions (Width: 57.5, Depth: 26.5, Height: 33) in inches		
7.	Internal Dimensions (Width: 52.2, Depth: 19.9, usable Height in main section: 25.5) in inches		
8.	Inner cabinet should be made of Pre-Painted steel		
9.	Wheels and Door Lockable option		
10.	Power Supply: 220V to 240V - 50HZ		
11.	Power Consumption: 210 Watts		
12.	PUF insulation of 60 mm thickness		
13.	The compressor should carry 3-years warranty		
14.	Should come with one basket.		

Additional eligibility criteria

Manufacturer should have a nationwide dealer network with strong technical support available locally and provide documentary evidence for the same.


Professor Vijay K. Chaudhary
निदेशक / Director
नवपरिवर्तन : सक्रामक रोग अनुसंधान, शिक्षा और प्रशिक्षण
Center for Innovation in Infectious Disease
Research, Education & Training (CIIDRET)
दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - 110021
University of Delhi South Campus, New Delhi-110021