

**Department of Genetics  
University of Delhi South Campus  
New Delhi-110021**

**E-Tender for Cooled Incubator (1 unit)**

**Description and Technical Specifications:**


The quotations are invited under two-bid system for supplying and installation of Cooled Incubator which is required to operate continuously at 20-22°C and as specifications described below:

1. Microprocessor controlled cooled incubator atleast 250 L validated usable space with temperature control from 0°C to 60°C and forced air circulation with 2 fans atleast for better temperature stability.
2. Temperature control range: 0°C to 60°C or better.
3. Temperature control accuracy: within  $\pm 0.1$  °C at 20°C
4. Temperature uniformity:  $\pm 0.1$ °C to  $\pm 0.5$ °C at 20°C.
5. Shelf: Atleast three adjustable shelf for incubation or storage
6. Alarm: Audible and visible alarm for temperature deviation more than 1°C from set point, and when timer operation has expired.
7. Control: simple programmable controller capable of holding 4-5 programs of not more than 10-15 steps each.
8. Refrigeration unit: The compressor must use CFC-free refrigerant.
9. Certification: ISO & CE certified or certified by any other internationally recognized Quality Management system.
10. Warranty: Atleast 24 months from the date of installation with facile after sales service.
11. It should have stainless steel inner body, powder coated steel (outer), inner tempered safety glass door and outer silicon packed magnetic door or better.
12. Machine should have proper insulation to prevent heat loss. It should have silent fan motor and automatic defrost function for better temperature uniformity and quiet operation.
13. It should have Back light LCD to display current value and set value simultaneously.

The vendor should enclose user list of minimum 10 users of the same machine supplied to various reputed institutes/research labs/ICAR/CSIR institutes throughout India in last 3 years.

**Eligibility Criteria**

1. Bidder should be manufacturer/authorized partner /reseller of the manufacturer and a letter of Authorization from manufacturer for the same and specific to the tender should be enclosed. The bidder should also be the Authorized Service provider.
2. As per recommendation of the purchase committee, vendor/bidder will be requested to give demonstration onsite (South Campus) for technical capabilities of the equipment quoted in the bid, failing which the bidder will be disqualified without any further communication.
3. Bidder should be financially sound to execute the order. Certificate to this effect should be issued by any nationalized/scheduled Bank showing its annual turnover of at least 25 lakhs each in the last three financial years.

  
डा. अरुणा नाओरम / Dr. Aruna Naorem  
सहायक प्राध्यापक / Asst. Professor  
आनुवंशिकी विभाग  
Department of Genetics  
दिल्ली विश्वविद्यालय दक्षिण परिसर  
University of Delhi South Campus  
नई दिल्ली - ११००२१ / New Delhi-110021

4. The bidder shall provide the Registration number of the firm along with the LST/CST/WCT No. and the PAN Number issued by the concerned authorities.
5. Separate technical and price bid should be uploaded: Technical Bid with Technical Compliance, all technical administration documents, original technical literature/catalogue (in support of the mentioned specifications & highlight the above features), Terms and Conditions of sales and Price Bids should contain the prices, terms of delivery, sales, payment terms, should be uploaded.
6. Copy of published literature in the form of catalogue/operating manual/website should be attached. Compliance sheet should mention the page number/website of the published catalogue/OM for each technical point.
7. The bidder must not be blacklisted by University of Delhi. A Certificate or undertaking to this effect must be submitted.
8. An undertaking stating the compliance of the conditions should be attached.

**No Commitment to Accept Lowest or Any Tender**

- University of Delhi shall be under no obligation to accept the lowest or any other offer received in response to this tender notice and shall be entitled to reject any or all offers.

**Price quotation**

- F.O.R. UDSC price to be quoted which includes custom duty and clearance charges if pricing is in foreign currency as well as transportation charges to the site of installation. CDEC will be provided.
- Payment will be made against successful installation. Advance payment will NOT be made.
- Price quotation must be valid for 6 months.

**Warranty /Comprehensive Maintenance**

- Machine should be warranted for minimum 24 months from the date of installation with facile after sales service.
- Installation and commissioning to be completed up to the satisfaction level of the technical purchase committee followed by training and handling by the bidders Price quotation.

**NOTE: Submission of hard copy of any of the documents mentioned above will not be accepted. All the relevant documents must be scanned and uploaded on the e-procurement portal.**

Published Date: 25-01-2019 (15:00 Hrs)


Bid Document Download Start Date: 28-01-2019 (09:00 Hrs)

Bid Submission Start Date: 29-01-2019 (09:00 Hrs)

Bid Submission End Date: 12-02-2019 (17:00 Hrs)

Bid Opening Date: 14-02-2019 (11:00 Hrs)

All documents related to tender can be found on CPPP portal

  
डॉ. अरुणा नाओरम / Dr. Aruna Naorem  
सहायक प्राध्यापक / Asst. Professor  
आनुवंशिकी विभाग  
Department of Genetics  
दिल्ली विश्वविद्यालय दक्षिण परिसर  
University of Delhi South Campus  
नई दिल्ली- ११००२१ / New Delhi- 110021