

**Department of Biochemistry
University of Delhi South Campus
Benito Juarez Road, New Delhi-110021**

Quotation invitation letter

Ref.No. UDSC/Biochem/2018-19/105

Dated: 11-02-2019

PURCHASE OF DOUBLE BEAM UV VIS SPECTROPHOTOMETER

Quotations are invited through e-procurement for the following equipment to be submitted online latest by 28TH February, 2019

Name of the equipment: Double Beam UV VIS Spectrophotometer

Specifications and Technical Details

ITEM	DESCRIPTION
Optical System	Double beam
Wavelength Range	190-1100 nm
Spectral Bandwidth	1nm
Wavelength Accuracy	±0.3 nm or better
Wavelength Repeatability	≤0.1nm or better
Photometric Accuracy	± 0.2 %T (0~100%T) or better
Photometric Repeatability	± 0.1%T (0~100%T) or better
Photometric Range	0 - 200 %T ; -3 to 3 Abs.
Stray Light	≤ 0.03 %T or better
Stability	± 0.0003 Abs./hr @ 500 nm or better
Baseline Flatness	± 0.001Abs or better
Operating Mode	%T, Abs., Concentration
Scanning Speed	Fast, Medium & Slow
Wavelength Setting	Auto
Display	Minimum 10" TFT Colored Touch Screen
Light Source	Deuterium Lamp & Tungsten Lamp
Detector	Silicon Photodiode
Output	USB Host / USB Drive / RS 232 / SD Card
Power	AC 220V/50Hz OR AC 110V/60Hz
Dimension	Not more than 600x500x250mm
Weight	20 kg or lower

Eligibility Criteria

- Bidder should be Manufacturer / Authorized Partner / Reseller of the manufacturer and a Letter of Authorization from manufacturer for the same and specific to the tender should also be enclosed.
- The Bidder shall provide the Registration number of the firm along with the LST/CST/WCT No. and the PAN Number issued by the concerned authorities.
- Separate technical and price bid should be uploaded.
- Technical Bid must contain a Technical Compliance sheet. Technical compliance sheet must have the page number of a published catalogue/technical specification/accessory information/manufacture website as proof for compliance with each technical point. Relevant copies of published catalogue / technical specifications / accessory information / manufacturer website that have been listed as proof should be attached
- The bidder must not be blacklisted by Delhi University. A Certificate or undertaking to this effect must be submitted.
- The bidder must provide buy back for the old **UV Visible Spectrophotometer (Model, Systronics, 2203)** and the discount offered must be mentioned in the financial bid.

Instructions to Suppliers

Two types of quotations are to be submitted in e-procurement site: Technical and Financial.


Technical Quotations should contain all details of technical specifications of the equipment including manufacture and model details along with commercial terms and conditions, warranty and delivery schedule. The quote should be valid for a minimum of 90 days from the due date. Pictures of the equipment must be included with similar specifications as above with logo or name of the company on them. Authorization certificates from the principal manufacturer, if any, must be provided. Proprietary certificate for any proprietary or patented item must be provided. Unique features may be highlighted along with their applications and advantages over other available accessories. A user list of at least 5 users of the equipment quoted (Who have installed and used them successfully for years) with addresses/ e-mail IDs/ Phone nos should also be provided.

Financial Quotations should contain price of the equipment, accessories, discount if any, packaging and forwarding charges, air freight and insurance charges, custom duty and clearance as well as transportation charges to the site of delivery/installation.

Payment will be made against successful delivery and after confirmation that the item is intact and functional.

No Commitment to Accept Lowest or Any Tender

- Demonstration of the item or its use will be required, if needed
- Institute reserves the right to order equipment with better quality and suitability over lower price and to accept or reject any or all quotations without assigning reasons thereof.
- University of Delhi shall be under no obligation to accept the lowest or any other offer received in response to this tender notice and shall be entitled to reject any or all offers. University of Delhi will not be obliged to meet and have discussions with any vendor, and or to listen to any representations.
- The supplier should provide quotation or tender through e-procurement site.

Head 
Department of Biochemistry
University of Delhi South Campus
New Delhi-110021