



**DEPARTMENT OF BIOCHEMISTRY
UNIVERSITY OF DELHI SOUTH CAMPUS**

Benito Juarez Road
New Delhi-110021, India
FAX: 91-11-24115270
Tel.: 91-11-24114157/58
E-mail: vkchaudhary@south.du.ac.in

2nd March 2015

Sub: " Supply of Multichannel pipets, Department of Biochemistry 280215 "

Dear Sir,

Please submit your lowest prices (INR/CBW/direct import) (as two part bid i.e. technical and financial) for the supply **Multichannel pipets**. **It is important that the variations, if any, are mentioned clearly in the technical bid in the compliance format.** The two bids should be kept in sealed envelopes clearly marked as **technical or financial** and these two sealed covers should be kept in another envelope marked as **"Offer for the supply Multichannel pipets"**. The order may be for the requested quantity or lesser (not less than 12 units total)

Supply of Multichannel pipets of the following specification:

A. 8-channel pipet 5-50 ul. with following specifications		Quantity required – Seven (7)
S.No.	Description	Compliance (Yes / No)
1.	Fully autoclavable multiple times without disassembly	
2.	Thumb-tip volume selection without changing hand position	
3.	Calibration without tools	
4.	Tip loading to be sealed using v-rings / o-rings	
5.	Manifold should rotate 360° in both directions	
6.	4-digit display	
7.	Accuracy + 1% or better at max setting	
8.	Coefficient of variation + 0.5% or better	
9.	Sub-division – 0.1 ul	
10.	Supplied along with 1-2 packs of tips and required tools	
11.	Comprehensive warranty for 1 years or more	
B. 8-channel pipet 20-200 ul. with following specifications		Quantity required – Seven (7)
S.No.	Description	Compliance (Yes / No)
1.	Fully autoclavable without disassembly	
2.	Thumb-tip volume selection without changing hand position	
3.	Calibration without tools	
4.	Tip loading to be sealed using v-rings / o-rings	
5.	Manifold should rotate 360° in both directions	
6.	4-digit display	
7.	Accuracy + 0.8 % or better or better at max setting	
8.	Coefficient of variation + 0.3% or better	
9.	Sub-division – 0.2 ul	
10.	Supplied along with 1-2 packs of tips and required tools	
11.	Comprehensive warranty for 1 years or more	

V.K. Chaudhary
Dr. V.K. Chaudhary
Professor & I
Department of Biochemistry
University of Delhi South Campus
New Delhi- 110021

Other conditions:


1. The supplier should be the original manufacturer or **authorized main/sole distributor** (attach sole distributorship certificate). The distributorship should be at least 5 years old and the tenderer must have sold and serviced similar equipments from the manufacturer in the Delhi academic institutions worth Rs.10 lacs / year or Rs. 15 lacs in the last three years.
2. The manufacturer should be a reputed company either a Government of India undertaking or should be multinational with presence in 5 countries and in India for the last 5 years.
3. The offer should be valid for a period of 3 months.

The offer in two separate envelopes marked 'Technical' and 'Financial', and then kept in one envelope marked "Tender for **Supply of Multichannel pipets**" should reach the undersigned by **12th March 2015**.

It is important to provide compliance to technical items in the tabular form as attached including the compliance to the point mentioned under 'Other Conditions'

Yours sincerely,

Prof Vijay K. Chaudhary
Professor
Department of Biochemistry,
First Floor, Engineering Dept Building
University of Delhi South Campus,
New Delhi 110021
INDIA



Dr. V.K. Chaudhary
Professor & ~~Head~~
Department of Biochemistry
University of Delhi South Campus
New Delhi - 110021