


Quotation for Ice Flaking Machine

Sealed quotations (technical and financial bids to be submitted separately) are invited from original equipment manufacturer or their authorised representatives for purchase of Ice Flaking Machine with specifications given below:

1. Machine should CE, VDE and GS approved and IS09001 certified
2. The unit should have at least production capacity of 120 Kg/24Hrs flakes ice with built-in highly hygienic PUF insulated bin storage capacity of 27 Kg
3. Compressor: Hermetically sealed
4. Refrigerant: R-134a CFC Free refrigerant gas
5. Condensation/Cooling: Air Cooled
6. Freezing Cylinder: Stainless Steel
7. Exterior (Chamber): Stainless Steel
8. Noise Level: Very Low Noise Level
9. Alarms Indications: Visual LED
10. Control: Microprocessor Control with LED indicator.
11. Operating Amps/Power Consumption: 2.0 Amps/480 Watts
12. Power Requirement: 220-240V /50Hz
13. Operating Temp. : 10 to 38 °C
14. Safety control: Microprocessor control against failure of refrigerant & water.
15. Machine operation should stop on non-availability of water and resumed when water is available.
16. Machine should stop when the Bin is full and resumes when sufficient ice is taken from the Bin.
17. An out let should be provided to drain water from the Bin to protect it from contamination.
18. Machine should stop and indicate if the Refrigerant is not sufficient to produce Ice.
19. Unit should have adjustable legs to keep the machine in level.
20. The unit should have 4 star rating with energy saving features
21. The participating vendor should have proven after sales service record and trained manpower.
22. Warranty: Minimum one year warranty should be included with the offer. (Cost of two years of additional warranty should be quoted as optional)
23. Prices should include transportation, customs clearance (if required) and on-site installation.
24. List of installations (preferably in the NCR region) should be provided.
25. On-site demonstration of the instrument may be required.

Last date for submission of quotation : **15th January 2016**


31/12/15
विभागाध्यक्ष / Head
आनुवंशिकी विभाग / Department of Genetics
दिल्ली विश्वविद्यालय दक्षिणी परिसर
University of Delhi South Campus
बेनिटो हुआरेज रोड / Benito Juarez Road
नई दिल्ली-110021