


Quotation for Colorimeter

Sealed quotations (technical and financial bids to be submitted separately) are invited from original equipment manufacturer or their authorised representatives for purchase of Colorimeter with specifications given below:

High Stability	Wavelength reproducibility
Reproducibility	+/-5nm or better
Filter	405,492,546,578,60nm; or 420,470,530,620,660nm (glass)
Wavelength Range	420-660nm
Photoelectric Range	T% : 0-100%, ABS:log T O-2
Photoelectric accuracy	2T or better
Light Source	6V 0.5 A tungsten lamp or better
Indicator	AE-11M 250U A Micro -ammeter, AE-11D three digital display or compatible
Power Supply	Indian standard electric requirement

- The participating vendor should have proven after sales service record and have trained manpower.
- Warranty: Minimum one year warranty should be included with the offer. (Cost of two years of additional warranty should be quoted as optional)
- Prices should include transportation, customs clearance (if required) and on-site installation.
- List of installations (preferably in the NCR region) should be provided.
- On-site demonstration of the instrument could be asked for.

Last date for submission of quotation : **15th January 2016**


31/12/15
विभागाध्यक्ष / Head
आनुवंशिकी विभाग / Department of Genetics
दिल्ली विश्वविद्यालय दक्षिणी परिसर
University of Delhi South Campus
बेनिटो हुआरेज रोड / Benito Juarez Road
नई दिल्ली-110029 / New Delhi-110021